

Acoustical Calibrator Test Report No.:11691

Calibrator: Norsonic Type 1253-250Hz

Serial no: 28326

Level: 124.07 dB

Frequency: 251.23 Hz

The stated level is valid at measurement conditions.

Calibrator signal distortion: 0.7 %

Frequency stability: 0.01 %

Measured according to IEC 60942.

The stated level is relative to 20 μ Pa.

Results are traceable to PTB, Germany or NIST, USA with a calculated uncertainty less than 0.10 dB (2 \times sd).

Scantek, Inc.

7060L Oakland Mills Road, Columbia MD 21046
Phone 410 290 7726 Service@ScantekInc.com

Reference conditions:

Pressure: 101.325 kPa

Temperature: 23.0 °C

Relative humidity: 50 %RH

Measurement conditions:

Pressure: 101.44 kPa

Temperature: 21.8 °C

Relative humidity: 39 % RH

Date: 2003-11-10

Signature: _____

Acoustical Calibrator Test Report No.:11691

Calibrator: Norsonic Type 1253-250Hz

Serial no: 28326

Level: 124.07 dB

Frequency: 251.23 Hz

The stated level is valid at measurement conditions.

Calibrator signal distortion: 0.7 %

Frequency stability: 0.01 %

Measured according to IEC 60942.

The stated level is relative to 20 μ Pa.

Results are traceable to PTB, Germany or NIST, USA with a calculated uncertainty less than 0.10 dB (2 \times sd).

Scantek, Inc.

7060L Oakland Mills Road, Columbia MD 21046
Phone 410 290 7726 Service@ScantekInc.com

Reference conditions:

Pressure: 101.325 kPa

Temperature: 23.0 °C

Relative humidity: 50 %RH

Measurement conditions:

Pressure: 101.44 kPa

Temperature: 21.8 °C

Relative humidity: 39 % RH

Date: 2003-11-10

Signature: _____

Comment: